

器具	クランプ メータ[A]	コンセント	CT出力電 圧[V]	実電力[W]	WHM0.4表 示[W]	CT値より 計算[W]	WHM0.4誤 差率	CT計算結 果の誤差率
		電圧[V]	圧[V]					
トースター	9.95	100.1	1.622	996	913	973	92%	98%
	9.85	100.1	1.610	986	907	966	92%	98%
	9.86	100.1	1.605	987	904	963	92%	98%
こたつ	5.31	101.8	0.878	541	492	527	91%	97%
	5.24	101.6	0.873	532	469	524	88%	98%
	5.23	101.6	0.871	531	480	523	90%	98%
ヒータ+電球	1.14	102.9	0.199	117	104	119	89%	102%
	1.14	102.8	0.199	117	104	119	89%	102%
	1.14	102.8	0.199	117	106	119	90%	102%
電球	0.48	103.2	0.092	50	48	55	97%	111%
	0.48	103.1	0.092	49	48	55	97%	111%
	0.48	103.0	0.092	49	49	55	99%	111%
半田ごて	0.17	103.4	0.040	18	19	24	108%	137%
	0.17	103.4	0.040	18	19	24	108%	137%
	0.17	103.3	0.040	18	19	24	108%	137%
ミニ電球	0.03	102.9	0.012	3	4	7	130%	240%
	0.02	102.9	0.012	2	4	7	194%	360%
	0.03	103.0	0.012	3	4	7	129%	240%

RL=490->500補正 係数6.0

